

CONSENT FORM FOR CARDIAC MONITOR TEST

Thank you for choosing our clinic for your cardiovascular care. To ensure the smooth progress of your arrhythmia monitor test, please take note of the following instructions:

Device Responsibility:

You are responsible for the monitor throughout its use. Promptly return the monitor to our office after the study ends to avoid delays in your results. If no follow-up appointment is scheduled, contact our office to arrange one.

Device Usage Tips:

- Remove the monitor for showers, leaving the electrodes in place.
- Change the adhesive patches every 3 days on average or sooner if edges pull away.
- Disregard lights on the WiFi side; they indicate transmission stages and do not affect data.

Battery Maintenance:

- Change the battery when orange lights appear by the battery icon or every 12 hours.
- One battery should always be in the charger and plugged in; the other in the monitor.

Monitoring Lights:

- No lights on the device at times are normal; the device is still functioning.
- Ignore orange lights on the WiFi symbol; data is still recorded and stored.

Troubleshooting and Additional Patches:

- For issues like a missing green light after placing the battery, contact us for troubleshooting.
- Visit the office for additional adhesive patches; no appointment or call required.

By signing below, you confirm your commitment to wearing the monitor for the prescribed time and agree to return the device promptly. Failure to comply may result in delays in receiving your results and potential charges for the device.

Patient Name: _____

Patient Signature: _____

Date: _____

This form will be scanned and added to your medical chart here at our office. If you would like a copy for your records please let us know so we can provide you with one.